



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 2333885

Master Contract: 211107

Project: 2333885

Date Issued: October 7, 2010

Issued to: Winters Instruments
121 Railside Rd
Toronto, ON M3A 1B2
Canada
Attention: David Reay

The products listed below are eligible to bear the CSA Mark shown



Sam Zaffino

Issued by: Sam Zaffino

PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

Pressure transmitter Models; LM6 Series; Type 6 enclosure, Rated Input: 12 to 36Vdc, 4-20mA, Pressure 0 to 300psi, ambient -10 to 70°C.

Pressure transmitter Models; LM7 Series; Type 6 enclosure, Rated Input: 12 to 36Vdc, 4-20mA, Pressure 0 to 300psi, ambient -10 to 70°C.

Pressure transmitter Models; LM3 Series; Type 6 enclosure, Rated Input: 12 to 36Vdc, 4-20mA, Pressure 0 to 300psi, ambient -10 to 70°C.

Pressure transmitter Models; LIM Series; Type 6 enclosure, Rated Input: 12 to 36Vdc (two wire version) or 14 to 36Vdc (three wire version), 4-20mA, Pressure 0 to 8700psi, ambient -25 to 85°C.

Pressure transmitter Models; LE3 Series, Type 6 enclosure, Rated Input: 8 to 32Vdc (two wire version) or 14 to 36Vdc (three wire version), 4-20mA, Pressure 0 to 5000psi, ambient -25 to 85°C.

Note: The units are rated SELV input, Pollution Degree 4 & Installation Category I

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements



CSA INTERNATIONAL

Supplement to Certificate of Compliance

Certificate: 2333885

Master Contract: 211107

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2333885	October 7, 2010	Evaluation of model series; LE3, LIM, LM3, LM6 & LM7 to C22.2 No. 61010-1-04