

Process Media and Material Compatibility: Since the wetted parts of the diaphragm seal assembly, which consist of the lower housing and diaphragm element itself, as well as any gaskets, will come in contact with the process, it is imperative to ensure the proper material is selected. The material selected must be compatible with all the various chemicals the seal may come in contact with while in service.

Temperature (Process & Ambient): Each diaphragm seal assembly is filled with an amount of fill fluid at ambient temperature. The increase of process or media temperature can cause the fill fluid to expand, creating a larger volume and pressure build up in the assembly.

Pressure Instrument: The diaphragm seal chosen must have the ability to provide enough volumetric displacement to drive the instrument through its full range. A rule of thumb is as follows: The lower the pressure, the more displacement is needed. The larger the gauge, the more displacement is needed. The smaller the diaphragm, the less displacement it can provide.

Process Connection and Orientation: The most common connections for diaphragm seals are threaded from 1/4" to 1" NPT and flanged. Be sure to specify both the flange size and rating when selecting a flanged seal.

Pressure Range: As stated above the diaphragm selected must be able to handle the pressure range to be sensed. The lower the pressure, the larger the diaphragm element must be to ensure sensitivity and proper displacement.

Response Time: Response time to sudden changes in process pressure can be effected by many things, such as size of the diaphragm, viscosity of the fill fluid, ambient and process temperatures and the use of capillary for remote mounted gauges.

Fill Fluid: Fill fluid selection is critical to the performance of the diaphragm seal assembly. Process and ambient temperatures, pressure and compatibility with the process media must be considered. Inert fills such as Halocarbon®, Fluorolube® and Krytox® must be specified when oxidizers are present.

Catalogue Page	Series	Instrument Connection			Process Connection				Wetted Parts Material								Pressure Range						RoHS compliant									
		1/4" NPT	1/2" NPT	1/2" NPT Male	1/4" NPT	1/2" NPT	3/4" NPT	≥1" NPT	Tri-Clamp®	RF Flange	Wafer Flange	Buna	Hastelloy® C	Monel®	PVC	Stainless Steel	Tantalum	PTFE	FKM	Other	Vac - 200 psi	30-600 psi		Vac - 1,000 psi	30-1,000 psi	30-1,300 psi	30-2,500 psi	50-5,800 psi	1,000-6,000 psi			
160	D10 PVC 2-Piece Seal	X			X	X								X			X				X											
161	D15 Threaded Flush Mount Seal	X					X								X													X	X	X		
162	D20 Sanitary Seal	X						X							X							X									X	
163	D30 High Pressure Seal	X			X										X															X	X	
164	D40 Medium Duty Seal	X	X		X	X									X									X							X	
165	D44 Flanged Seal (PTFE)	X	X						X								X						X									
167	D44 Flanged Seal All Welded	X	X						X			X	X	X	X	X	X	X	X							X						
169	D46 Flanged Flush Mount Seal	X	X						X					X										X								
170	D70 Heavy Duty 2-Piece Industrial Seal	X	X		X	X	X	X				X	X	X	X		X	X								X					X	
172	D71 NACE 2-Piece Seals	X	X		X	X	X					X	X				X	X								X						
174	D81 Isolation Ring	X	X	X				X		X	X	X		X	X	X	X	X	X				X									
176	D85 In Line Flow Thru Seal	X	X		X	X	X	X						X								X										
177	D90 Saddle Weld Seal	X	X				X							X	X	X	X	X	X			X										

Hastelloy® is a registered trademark of Haynes International, Inc.
 Fluorolube® is a registered trademark of Gabriel Performance Products, LLC.
 Halocarbon® is a registered trademark of Halocarbon Products Corporation
 Krytox® is a registered trademark of DuPont Corporation

Monel® is a registered trademark of Inco Alloys International
 Tri-Clamp® is a registered trademark of Tri-Clover

Note: When selecting a diaphragm seal, always refer to ASME B40.100 (2013)